

## **ΠΑΡΑΡΤΗΜΑ 1**


## Calibration Certificate


Narda Safety Test Solutions hereby certifies that the referenced equipment has been calibrated by qualified personnel to Narda's approved procedures. The calibration was carried out within a certified quality management system conforming to ISO 9001:2000.

The metrological confirmation system for test equipment complies with ISO 10012-1.

Object	Antenna, Three-Axis, E-Field, 27 MHz to 3 GHz
Part Number (P/N)	3501/03
Serial Number (S/N)	K-0105
Manufacturer	Narda Safety Test Solutions
Customer	
Date of Calibration	22-Jan-2010
Results of Calibration	Test results within specifications
Confirmation interval recommended	24 Months
Ambient conditions	23°C +/-3°C (20...60)% rel. humidity
Calibration procedure	3000-8702-00A

Pfullingen, 22-Jan-2010

  
Person in charge  
Geyer

  
Head of Laboratory  
N. Moll



This certificate may only be published in full, unless permission for the publication of an approved extract has been obtained in writing from the Managing Director.

Certified by DQS according to  
DIN EN ISO 9001:2000  
(Reg.-No. 99379-QM)

## Calibration Certificate

Narda Safety Test Solutions hereby certifies that the object referred to in this certificate has been calibrated by qualified personnel using Narda's approved procedures. The calibration was carried out in accordance with a certified quality management system which conforms to ISO 9001

OBJECT	Selective Radiation Meter, Basic Unit, SRM-3006
MANUFACTURER	Narda Safety Test Solutions GmbH
PART NUMBER (P/N)	3006/01
SERIAL NUMBER (S/N)	B-0065
CUSTOMER	
CALIBRATION DATE	2010-01-14
RESULT ASSESSMENT	within specifications
AMBIENT CONDITIONS	Temperature: (23 ± 3)°C Relative humidity: (25 to 75)%
CALIBRATION PROCEDURE	3006-8701-00A

ISSUE DATE: 2010-01-15

  
CALIBRATED BY:  
Paul Geyer

  
AUTHORIZED SIGNATORY:  
Norbert Moll

This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory. Calibration certificates without signature are not valid.

MANAGEMENT  
SYSTEM  
  
Certified by DQS against  
DIN EN ISO 9001:2000  
(Reg.-No. 099379)

## Calibration Certificate

Narda Safety Test Solutions hereby certifies that the object referenced to this certificate has been calibrated by qualified personnel using Narda's approved procedures. The calibration was carried out in accordance with a certified quality management system which conformed to ISO 9001.

OBJECT	RF-cable SRM, 9 kHz to 6 GHz, N 50 ohm
MANUFACTURER	Narda Safety Test Solutions GmbH
PART NUMBER (P/N)	3602/01
SERIAL NUMBER (S/N)	AA-0109
CUSTOMER	
CALIBRATION DATE	2010-01-20
RESULT ASSESSMENT	Within specifications
AMBIENT CONDITIONS	Temperature: (23 ± 3) °C Relative humidity: (20 to 60) %
CALIBRATION PROCEDURE	3000-8703-00A

ISSUE DATE: 2010-01-20

  
\_\_\_\_\_  
CHECKED BY  
Geyer

  
\_\_\_\_\_  
AUTHORIZED SIGNATORY  
N. Moll

This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory. Calibration certificates without signature are not valid.



Certified by DQS according to  
DIN EN ISO 9001:2000  
(Reg.-No. 99379-QM)